



# ALTA DIAGNOSTICS, INC.

2560 BUSINESS PARKWAY STE C, MINDEN, NV 89423

(800) 359-9691 (775) 267-3001 FAX: (775) 267-1142

## LIQUID HUMAN URINE CONTROL

LEVEL 1 LOT#185931

LEVEL 2 LOT#185932

EXPIRATION DATE: June 2011

18 MONTH OPEN VIAL STABILITY

# UA

### INSTRUMENTS

You may enter the mean and range values for your instrument in these last columns. ↓ ↓

ANALYTES		TARGET VALUES		BECKMAN		DIMENSION		HITACHI		VITROS			
		LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2
		<b>Amylase</b>	U/L			40 ± 10	140 ± 30	15 ± 5	50 ± 10	208 ± 20	751 ± 70	50 ± 15	160 ± 30
<b>Calcium</b>	MG/DL	7	4	5.6 ± 2	3.1 ± 1.5	5.7 ± 2	3.1 ± 1.0	5.6 ± 2.0	3.1 ± 1.5	5.7 ± 2	3.3 ± 1.5		
<b>Chloride</b>	MEQ/DL	80	220	70 ± 14	210 ± 20	80 ± 15	220 ± 20	70 ± 14	211 ± 20				
<b>Creatinine</b>	MG/DL	20	60	26 ± 4	74 ± 12	21 ± 4	66 ± 10	21 ± 4	65 ± 12	22 ± 4	68 ± 12		
<b>Glucose</b>	MG/DL	20	80	22 ± 4	81 ± 9	20 ± 4	80 ± 8	21 ± 4	82 ± 8	21 ± 4	85 ± 9		
<b>Magnesium</b>	MG/DL	10	3	12 ± 4	3.2 ± 1.5	8.0 ± 2	3.1 ± 1.0	8.1 ± 4	3.2 ± 1.5	8.2 ± 4	3.2 ± 1.0		
<b>Phosphorus</b>	MG/DL	7	20	6.1 ± 3	21.1 ± 4	6.2 ± 3	20 ± 4	6.9 ± 3	20 ± 5	5.6 ± 3	21 ± 5		
<b>Potassium</b>	MEQ/DL	15	50	15 ± 3	64 ± 5	15 ± 3	63 ± 5	15 ± 3	62 ± 5	15 ± 3	63 ± 5		
<b>Sodium</b>	MEQ/DL	55	170	58 ± 12	194 ± 15	60 ± 12	194 ± 15	57 ± 12	193 ± 15	57 ± 12	195 ± 20		
<b>Tot-Protein</b>	MG/DL	12	70	21 ± 4	75 ± 12	22 ± 4	77 ± 12	20 ± 4	74 ± 12	22 ± 6	70 ± 12		
<b>Urea/Bun</b>	MG/DL	80	330	95 ± 12	310 ± 30	80 ± 12	320 ± 30	80 ± 13	320 ± 30	80 ± 15	320 ± 40		
<b>Uric Acid</b>	MG/DL	5	11	5.3 ± 2	10.1 ± 2	5.5 ± 2	10.5 ± 2	5.2 ± 2.0	10.3 ± 2	5.3 ± 2	9 ± 3		
<b>Albumin</b>	MG/DL	10	40	7 ± 2	50 ± 6			7 ± 2					

#### PRECISION

#### ADVANCE INSTRUMENTS

#### FISKE

#### WESCOR

LEVEL 1

LEVEL 2

LEVEL 1

LEVEL 2

LEVEL 1

LEVEL 2

LEVEL 1

LEVEL 2

<b>Osmolality</b>	m0sm/kg	195 ± 15	625 ± 35	195 ± 15	625 ± 35	195 ± 15	625 ± 35	195 ± 15	625 ± 35
-------------------	---------	----------	----------	----------	----------	----------	----------	----------	----------

10/1/10