

ALTA DIAGNOSTICS, INC.

18 MONTH

OPEN VIAL

STABILITY

3123 Research Way Ste 214, Carson City NV 89706 (800) 359-9691 (775) 283-5780 FAX: (775) 283-5787 LIQUID URINE CONTROL

FOR

MICROSCOPIC & HIGH SPECIFIC GRAVITY

	LOT #036819 POSITIVE Exp 02/20		LOT #036820 NEGATIVE Exp 02/20		PROCEDURE
	MICROSCOPIC	SPECIFIC GRAVITY	MICROSCOPIC	SPECIFIC GRAVITY	
HIGH	35	1.036		1.036	1. Shake well before using to assure
MEAN	20 CELL/HP ± 15	1.031 ± .005	0 CELL/HP	1.031 ± .005	complete mixing of the contents.
LOW	5	1.026		1.026	2. Remove bottle cap and pour 12 ml into a clean, dry conical centrifuge tube.*
DAY 1					
DAY 2					3. Centrifuge for 5 minutes at 2000 rpm. (A lower rpm may be used if this is called for in your laboratory procedure. However, a somewhat lower mean may result!)
DAY 3					
DAY 4					
DAY 5					4. Remove control from the centrifuge and at this time, if desired, take and record the specific gravity reading by placing a small urinometer in the centrifuge tube or, alternatively, transfer a few drops of the supernate to a refractometer.
DAY 6					
DAY 7					
DAY 8					
DAY 9 DAY 10					5. Pour off and discard all but 0.5 ml of the
DAT 10 DAY 11					supernate.
DAY 12					 Resuspend the sediment in the remaining 5 ml of supernate by touching the bottom of the tube to a vortex machine or by flicking the
DAY 12					
DAY 14					bottom of the tube with your finger.
DAY 15					7. Transfer a drop of the resuspended sediment to a clean dry microscope slide and
DAY 16					cover with a cover slip.
DAY 17					8. Count and record the average number of
DAY 18					cells found in 10 high power fields.
DAY 19					9. At the end of the month, add the column of
DAY 20					entries for MICROSCOPIC and/or SPECIFIC GRAVITY and enter the TOTAL at the bottom of the column. Determine the MEAN by
DAY 21					
DAY 22					dividing the TOTAL by the number of days the test was run.
DAY 23					10. Store at 2º - 8ºC. May be stored at room
DAY 24					temperature once bottle is in use.
DAY 25					*NOTE:The value range for Alta's Microscopic
DAY 26					Control is based on the parameters set forth in the above procedure. Laboratories using a procedure
DAY 27					with different parameters (i.e. volume, rpm and time
DAY 28					of centrifugation and amount of supernate discarded) should develop their own range of values
DAY 29					and mean for the control using their procedure.
DAY 30					
DAY 31					
TOTAL					
MEAN					