

ALTA DIAGNOSTICS, INC.

2560 BUSINESS PARKWAY STE C, MINDEN, NV 89423 (800) 359-9691 (775) 267-3001 FAX: (775) 267-1142 18 MONTH Open Vial Stability

LIQUID URINE CONTROL

FOR

MICROSCOPIC & HIGH SPECIFIC GRAVITY

LOT # 097019 Exp 3/12	MICROSCOPIC	SPECIFIC GRAVITY	PROCEDURE
HIGH	35	1.032	1. Shake well before using to assure complete
MEAN	20 CELL/HP ± 15	1.027 ± .005	mixing of the contents.
LOW	5	1.022	2. Remove bottle cap and pour 12 ml into a
DAY 1			clean, dry conical centrifuge tube.*
DAY 2			3. Centrifuge for 5 minutes at 2000 rpm. (A
DAY 3			lower rpm may be used if this is called for in your laboratory procedure. However, a
DAY 4 DAY 5			somewhat lower mean may result!)
DAY 6			4. Remove control from the centrifuge and at
DAY 7			this time, if desired, take and record the
DAY 8			specific gravity reading by placing a small urinometer in the centrifuge tube or,
DAY 9			alternatively, transfer a few drops of the
DAY 10			supernate to a refractometer.
DAY 11			5. Pour off and discard all but 0.5 ml of the
DAY 12			supernate.
DAY 13			6. Resuspend the sediment in the remaining
DAY 14			0.5 ml of supernate by touching the bottom of the tube to a vortex machine or by flicking the
DAY 15			bottom of the tube with your finger.
DAY 16			7. Transfer a drop of the resuspended
DAY 17			sediment to a clean dry microscope slide and
DAY 18			cover with a cover slip.
DAY 19			8. Count and record the average number of
DAY 20			cells found in 10 high power fields.
DAY 21			9. At the end of the month, add the column of
DAY 22			entries for MICROSCOPIC and/or SPECIFIC GRAVITY and enter the TOTAL at the bottom of the column. Determine the MEAN by dividing the TOTAL by the number of days the test was run.
DAY 23			
DAY 24			
DAY 25			
DAY 26			10. Store at 2º - 8ºC. May be stored at room
DAY 27			temperature once bottle is in use.
DAY 28			*NOTE:The value range for Alta's Microscopic
DAY 29			Control is based on the parameters set forth in the above procedure. Laboratories using a procedure with different parameters (i.e. volume, rpm and time of centrifugation and amount of supernate discarded) should develop their own range of values and mean for the control using their procedure.
DAY 30			
DAY 31			
TOTAL			
MEAN			procedure.