



ALTA DIAGNOSTICS, INC.

2560 BUSINESS PARKWAY STE C, MINDEN, NV 89423
 (800) 359-9691 (775) 267-3001 FAX: (775) 267-1142

LIQUID HUMAN URINE CONTROL

LEVEL 1 LOT#179701A LEVEL 2 LOT#179702A EXPIRATION DATE: JULY 2010
 18 MONTH OPEN VIAL STABILITY

UA

INSTRUMENTS

You may enter the mean and range values ↓

ANAL YTES		TARGET VALUES		BECKMAN		DIMENSION		HITACHI		VITROS		LEVEL 1	LEVEL 2	
		LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	
	Amylase	U/L			36 ± 9	128 ± 30	14 ± 4	47 ± 8	185 ± 20	655 ± 60	50 ± 15	155 ± 30		
	Calcium	MG/DL	7	4	5.9 ± 2	3.0 ± 1.5	6.1 ± 2	3.1 ± 1.0	5.9 ± 2	3.0 ± 1.5	6.1 ± 2	3.8 ± 1.5		
	Chloride	MEQ/DL	80	220	65 ± 14	205 ± 20	75 ± 10	217 ± 20	65 ± 14	204 ± 20				
	Creatinine	MG/DL	20	60	28 ± 4	70 ± 12	22 ± 4	60 ± 10	23 ± 4	60 ± 12	26 ± 4	63 ± 12		
	Glucose	MG/DL	20	80	22 ± 4	93 ± 9	20 ± 4	90 ± 9	21 ± 4	93 ± 9	20 ± 4	96 ± 9		
	Magnesium	MG/DL	10	3	13.0 ± 4	3.5 ± 1.5	12 ± 2	3.3 ± 1.0	13 ± 4	3.5 ± 1.5	13.1 ± 3	3.3 ± 1.0		
	Phosphorus	MG/DL	7	20	7.6 ± 3	23 ± 5	8.0 ± 2	22 ± 5	8.5 ± 3	23 ± 5	6.8 ± 3	22 ± 5		
	Potassium	MEQ/DL	15	50	15 ± 3	54 ± 5	15 ± 3	52 ± 5	15 ± 3	54 ± 5	17 ± 3	58 ± 5		
	Sodium	MEQ/DL	55	170	60 ± 12	174 ± 15	62 ± 12	170 ± 15	58 ± 12	174 ± 15	58 ± 12	173 ± 20		
	Tot-Protein	MG/DL	12	70	39 ± 10	97 ± 15	35 ± 7	90 ± 12	35 ± 8	95 ± 12	49 ± 6	85 ± 8		
	Urea/Bun	MG/DL	80	330	94 ± 12	340 ± 30	93 ± 12	359 ± 30	94 ± 12	360 ± 30	95 ± 15	330 ± 40		
	Uric Acid	MG/DL	5	11	6.4 ± 2	7.5 ± 2	6.5 ± 2	11.5 ± 2	6.4 ± 2	11.5 ± 2	6.3 ± 2	6.9 ± 2		
	Albumin	MG/DL	10	40	12 ± 3	39 ± 6								

		PRECISION		ADVANCE INSTRUMENTS		FISKE		WESCOR	
		LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2	LEVEL 1	LEVEL 2

Osmolality	mOsm/kg	210 ± 20	590 ± 50	210 ± 20	590 ± 50	210 ± 20	590 ± 50	210 ± 20	590 ± 50
-------------------	----------------	----------	----------	----------	----------	----------	----------	----------	----------